



# CALIBRATION REPORT

## PI TAPE®

Report Number:	SAMPLE
Date Issued:	07/07/07
Pi Tape Corporation Serial Number:	SAMPLE
Tape Size:	12" – 24" O.D.

### About this Calibration

Temperature	68 Degrees F ± 2 Degrees	
Relative Humidity	45% ± 15%	
Tension	Five Pounds	
Tape Tolerance	Within .001"	Within Mfg Tolerance of .001"
Uncertainty of Measurement	±.00025 Inches	

Calibration Performed by: Jerry Mathis

### Actual Findings and Procedure Numbers

Tape Reads	20.0000	Over 20.0000"	Procedure 1
------------	---------	---------------	-------------

Remarks: New Gage

### Calibration Equipment Used

Gage	Serial Number	Test Date	Due Date	NIST Traceable Report Number
20"	042199	03/12/07	03/12/08	72994-4

This tape has been checked at one or more points over a calibrated measuring standard. Calibration performed in accordance with ANSI/NC SL Z540-1-1994, ISO/IEC 17025 and the former MIL-STD 45662A.

**All standards are traceable to the National Institute of Standards and Technology**

**Jerry Mathis, Quality Assurance Manager**

This calibration report shall not be reproduced, except in full, without the written approval of the issuing laboratory. Results listed relate only to the item being calibrated.

### Pi Tape Corporation

344 N. Vinewood St. Escondido, CA 92029 USA  
Phone (760) 746-9830 Fax (760) 746-9196  
www.pitape.com