



# CALIBRATION REPORT

## PI TAPE®

### SAMPLE – SAMPLE – SAMPLE

Report Number:	052611
Date Issued:	May-26-11
Pi Tape Corporation Serial Number:	
Tape Size:	50 – 300mm O.D.

#### About this Calibration

Temperature	68 Degrees F $\pm$ 2 Degrees	
Relative Humidity	45% $\pm$ 15%	
Tension	2.25 kg.	
Tape Tolerance	Within .03mm	Within Mfg Tolerance of .03mm
Uncertainty of Measurement	$\pm$ .008mm	

Calibration Performed by: Jerry Mathis

#### Actual Findings and Procedure Numbers

Tape Reads		Over 125.00mm	Procedure 1
------------	--	---------------	-------------

Remarks: New Gage

#### Calibration Equipment Used

Gage	Serial Number	Test Date	Due Date	NIST Traceable Report Number
125mm	102098	May-12-11	May-12-12	86903-1

This tape has been checked at one or more points over a calibrated measuring standard. Calibration performed in accordance with ANSI/NC SL Z540-1-1994, ISO/IEC 17025 and the former MIL-STD 45662A.

**All standards are traceable to the National Institute of Standards and Technology**

Jerry Mathis, Quality Assurance Manager

This calibration report shall not be reproduced, except in full, without the written approval of the issuing laboratory. Results listed relate only to the item being calibrated.

#### Pi Tape Corporation

344 N. Vinewood St. Escondido, CA 92029 USA  
Phone (760) 746-9830 Fax (760) 746-9196  
www.pitape.com